JiangSu YuYue Medical Equipment & Supply CO., LTD.

Section 7

Declaration of Conformity

No.:	**YY-JS.XY-1007**
Edition/revision:	**A/0**
Status of control:	**Get control**
Issuing date:	**2024.05.17**
Prepared by:	Shufei
Reviewed by:	Hongxia. Mao
Approved by:	Guomin. jing

EU Declaration of Conformity

Manufacturer:JiangSu YuYue Medical Equipment & Supply CO., LTD.
Address: No.1 Baisheng Road Development Zone,Danyang,Jiangsu 212300 CHINA
SRN: CN-MF-000012834

EuropeanRepresentative: Metrax GmbH
Address: Rheinwaldstr. 22, 78628 Rottweil, Germany
SRN: DE-AR-000005481

Product Name： Aneroid sphygmomanometers
Model: XB-11, YM100, 3.02.008
Intended purpose: The product is used for measuring blood pressure.
Basic UDI-DI: 693325792233HV
EMDN Code: C9006
Classification: Ⅰm based on on MDR 2017/745 annex VIII rule 1
Conformity Assessment Route: MDR Annex IX chapter I, chapter III and Section 4

We declare that above mentioned products meet the provision of the following EC Council Directive and Regulation. All the supporting documents and files are retained under the premises of the manufacture. We are exclusively responsible for the Declaration of Conformity. There is no applicable common specifications for Aneroid sphygmomanometers. General applicable directives and regulation:
Medical Device Regulation (EU) 2017/745

Notified Body: TÜV SÜD Product Service GmbH, Ridlerstr. 65, 80339München, Germany

Notified Body Number : 0123

(EC) Certificate(s):G21 109546 0011 Rev.00

Expire date of the Certificate: 2029.05.09

Signature: Jie Mei

Name:Jie Mei

Position:Person Responsible for Regulatory Compliance

Place, Date of Issue: DanYang, Jiangsu , P.R.CHINA 2024-05-17